

Cliquid MD 3.4 HotFix 2 Release Notes



Introduction

For information about a previous software version, refer to the *Release Notes* that came with that software version.

New in Cliquid MD 3.4 HotFix 2

Enhancements and Fixed Issues

This HotFix includes the fix to the following issue:

- If a customized test was created with the IS column hidden and left empty, and a sample list was imported, then the values of the Internal Standard (IS) concentrations were incorrectly derived from the value in the Analyte Concentration column. This resulted in values of the concentrations being shown in the wrong columns. (CL-1564)

The HotFix also includes the fixed issue included in the following HotFix.

HotFix 1

- The version number has been updated on the Login dialog.

Known Issues and Limitations

- N/A

Installation

Install the HotFix

| Prerequisites |
|--|
| <ul style="list-style-type: none">The Cliquid MD 3.4 software is installed. |
| <p>Note: This HotFix is compatible with Cliquid MD 3.4 HotFix 1, and can be installed on computers on which this HotFixes has been installed. The previous HotFix is removed silently during installation of this HotFix.</p> |

1. Log on to the computer as a user with Administrator privileges.
2. Stop any acquisitions that are in progress and then deactivate the hardware profile.
3. Close all applications.
4. Download **Cliquid MD 3.4 HotFix 2** from [sciex.com/software-support/software-downloads](https://www.sciex.com/software-support/software-downloads).

Tip! To prevent potential installation issues, save the file to a location other than the computer desktop and disconnect any external USB storage devices before starting the installation.

5. After the download is complete, right-click the **CliquidMD3.4HF2.zip** file.
6. Click **Extract All**, select the file destination folder, and then click **Extract**.
7. After the extraction is complete, browse to the extracted folder and double-click the **CliquidMD_3.4_HotFix2.exe** file.
8. Follow the on-screen instructions to complete the installation.
9. After the installation is complete, do this to make sure that the HotFix is installed:
 - a. Open Control Panel and then click **Programs and Features**.
 - b. Make sure that **Cliquid MD 3.4 HotFix 2** is shown in the list of programs.
 - c. Close Control Panel.
10. Open the Analyst MD software and then activate the hardware profile. Refer to the documentation for the Analyst MD software.

Remove the HotFix

Remove the HotFix before removing the Cliiquid MD software.

1. Log on to the computer as a user with Administrator privileges.
2. Deactivate the hardware profile in the Analyst MD software.
3. Close all applications.
4. Open Control Panel and then click **Programs and Features**.
5. Select **Cliiquid MD 3.4 HotFix 2** and then click **Uninstall**.
6. Follow the on-screen instructions.

The HotFix is removed from the program list, and any files that were changed by the HotFix are reverted to the original Cliiquid MD 3.4 software version.

Note: If Cliiquid MD 3.4 HotFix 1 was removed during installation of the Cliiquid MD 3.4 HotFix 2, then it is not restored. It must be installed again after HotFix 2 is removed.

7. Close Control Panel.
8. Open the Analyst MD software and then activate the hardware profile. Refer to the documentation for the Analyst MD software.

Updated Files

The HotFix updates the following files in the C:\Program Files (x86)\AB SCIEX\cliiquid\Client\ folder:

- Cliiquid.exe
- Cliiquid.Resources.dll

Contact Us

Customer Training

- In North America: NA.CustomerTraining@sciex.com
- In Europe: Europe.CustomerTraining@sciex.com
- Outside the EU and North America, visit sciex.com/education for contact information.

Online Learning Center

- [SCIEX Now Learning Hub](#)

SCIEX Support

SCIEX and its representatives maintain a staff of fully-trained service and technical specialists located throughout the world. They can answer questions about the system or any technical issues that might arise. For more information, visit the SCIEX website at sciex.com or contact us in one of the following ways:

- sciex.com/contact-us
- sciex.com/request-support

CyberSecurity

For the latest guidance on cybersecurity for SCIEX products, visit sciex.com/productsecurity.

Documentation

This version of the document supercedes all previous versions of this document.

To view this document electronically, Adobe Acrobat Reader is required. To download the latest version, go to <https://get.adobe.com/reader>.

To find software product documentation, refer to the release notes or software installation guide that comes with the software.

Note: To request a free, printed version of this document, contact sciex.com/contact-us.

This document is provided to customers who have purchased SCIEX equipment to use in the operation of such SCIEX equipment. This document is copyright protected and any reproduction of this document or any part of this document is strictly prohibited, except as SCIEX may authorize in writing.

IVD

Software that may be described in this document is furnished under a license agreement. It is against the law to copy, modify, or distribute the software on any medium, except as specifically allowed in the license agreement. Furthermore, the license agreement may prohibit the software from being disassembled, reverse engineered, or decompiled for any purpose. Warranties are as stated therein.

Portions of this document may make reference to other manufacturers and/or their products, which may contain parts whose names are registered as trademarks and/or function as trademarks of their respective owners. Any such use is intended only to designate those manufacturers' products as supplied by SCIEX for incorporation into its equipment and does not imply any right and/or license to use or permit others to use such manufacturers' and/or their product names as trademarks.

CE

SCIEX warranties are limited to those express warranties provided at the time of sale or license of its products and are the sole and exclusive representations, warranties, and obligations of SCIEX. SCIEX makes no other warranty of any kind whatsoever, expressed or implied, including without limitation, warranties of merchantability or fitness for a particular purpose, whether arising from a statute or otherwise in law or from a course of dealing or usage of trade, all of which are expressly disclaimed, and assumes no responsibility or contingent liability, including indirect or consequential damages, for any use by the purchaser or for any adverse circumstances arising therefrom.

(GEN-IDV-09-10816-D)

For In Vitro Diagnostic Use. Product(s) not available in all countries. For more information contact your local sales representative or refer to sciex.com/diagnostics.

Rx Only.

Trademarks and/or registered trademarks mentioned herein, including associated logos, are the property of AB Sciex Pte. Ltd., or their respective owners, in the United States and/or certain other countries (see sciex.com/trademarks).

AB SCIEX™ is being used under license.

© 2021 DH Tech. Dev. Pte. Ltd.



AB Sciex Netherlands B.V.
1e Tochtweg 11,
2913LN Nieuwerkerk aan den IJssel
Netherlands

Cliquid MD 3.4 HotFix 2 Release Notes



AB Sciex Pte. Ltd.
Blk33, #04-06 Marsiling Industrial Estate Road 3
Woodlands Central Industrial Estate, Singapore 739256